Instructions for Use

SANIBEL MPT-II Printer





Item No. D-0100258 - 08/2015

1. Cautions

- RISK OF EXPLOSION IF BATTERY IS REPLACED BY AN INCORRECT TYPE
- DISPOSE OF USED BATTERIES ACCORDING TO LOCAL REGULATIONS
- DO NOT PLACE THE BATTERY IN FIRE OR HEAT THE BATTERY
- DO NOT CONNECT THE POSITIVE TERMINAL AND THE NEGATIVE TERMINAL OF THE BATTERY TO EACH OTHER WITH ANY METAL OBJECT
- DO NOT PLACE THE BATTERIES IN MICROWAVE OVENS, HIGH-PRESSURE CONTAINERS, OR ON INDUCTION COOKWARE
- IN THE EVENT THAT THE BATTERY LEAKS AND THE FLUID GETS INTO ONE'S EYE, DO NOT RUB THE EYE. RINSE WELL WITH WATER AND IMMEDIATELY SEEK MEDICAL CARE. IF LEFT UNTREATED THE BATTERY FLUID COULD CAUSE DAMAGE TO THE EYE
- THIS EQUIPMENT IS FOR PROFESSIONAL USE ONLY
- DO NOT OPEN THE COVER WHILE PRINTING. THE PRINTER HEAD MIGHT BE HOT
- DO NOT EXPOSE THE PRINTER TO ANY LIQUID.
- DO NOT DISASSEMBLE THE PRINTER
- IF PRINTER IS STORED FOR A LONGER PERIOD OF TIME, STORE THE BATTERY SEPARATELY
- FOR CORRECT BATTERY CHARGING MAKE SURE TO FOLLOW CHARGING INSTRUCTION

2. General Information

MPT-II printer kit includes:

- ☐ MPT-II rechargeable battery pack
- ☐ 40mm diameter thermal paper
- Power supply, medically approved 12V UE15WCP1-120125SPA
- ☐ Adapters for power supply for world-wide connection
- □ User manual

Connection cable supplied separately, depending on instrument.

3. Operation

Paper loading:

Open the lid by pushing on the sides, insert paper roll as shown, and close the lid.







Reorder paper from www.sanibelsupply.com or from your local distributor.

Battery pack insertion:

Insert battery as shown



Power on:

Push POWER BUTTON for two seconds to power ON or OFF.

One short beep will be heard at power ON, two short beeps at power OFF.

Green Power Indicator will be lit if printer is powered by battery.



Indicators:

Green • LED indicator	Blue LED indicator	Status	Sound	Note
Off O	Fast flash 🔵	Charging		Power On
Off O	On	Charging		Power Off
Off O	Slow flash	Battery nearly discharged		
Off O	On	Charging completed		Power On
Off O	Off	Charging completed		Power Off
On •	Off	Power ON, battery powered		
Slow flash	Slow flash	Out of paper	Веер	
Slow flash	Off O	Sleep mode		

Self-test:

While printer is powered OFF, press and hold PAPER FEED



button, then press and hold POWER BUTTON simultaneously. When beep is heard after approx. 3 seconds, release both buttons. and a test page will print with information on current status and character samples.

Paper feed:

When powered press PAPER FEED button. Paper will feed as long as the button is pressed.

Bluetooth setup:

Pairing with the instrument is necessary before printing via Bluetooth is possible.

The pairing process is done like this:

- 1. Power on printer
- 2. Let the instrument / PC search for other Bluetooth devices.
- A device called MPT-II should occur in the list of devices.
- 4. When prompted enter password "1234"
- 5. Finish pairing

For detailed instructions how to initiate Bluetooth search, please refer to your device's user manual.

Note: Do not have several printers powered on and within range while searching.

It is possible to pair up to 8 instruments with one printer. If more instruments are paired the oldest will be deleted.

4. Technical Specifications

Print mode:

Thermal line dot print Printing width: 48mm

Resolution: 8 dots/mm (203dpi)

Dots per line: 384 dots

Thermal paper:

Paper width = 56mm +/- 1mm max 40mm diameter

Printer mechanism life:

10^8 print pulses or 50km paper

Interface options: USB (Mini-USB)

RS-232 (6-pin mini-DIN)

Infrared (RAW_IR, VIR, IrDA)

Bluetooth Class2 (up to 10m range) – serial port profile (SPP)








RS-232 pinout:

- 1. TXD
- 2. RTS
- 3. DND
- 4. Reserved pins
- 5. RXD
- 6. Reserved pins



Battery pack:

2-cell Li-Ion battery pack 7.4V-1500mAh

Power supply / charger:

12V/1A UE15WCP1-120125SPA Maximum current consumption 0.5A



Physical properties:

Size: 102x75x45mm (WxLxH)

Weight: 205g including battery, without paper

Environmental specifications:

Operational temperature range: -10C to 50C (14F to 122F)

Operational humidity range: 20% to 85%

Storage temperature range: -20C to 70C (-4F to 158F)