

pH standard solutions, IUPAC, certified

Leverantör: Hach Lange









Traceable to IUPAC pH scale and NIST

For calibration of pH meters

Accuracy at 25 °C: $\pm 0,010$ or $\pm 0,05$ (pH 12,45)

Shelf life: 48 months from production date, except 24 months for pH 12.45
Observed shelf life after opening, following recommended storage and use: 3 months



VWR art.nr	Antal	pH-värde	pH accuracy	Reference temperature	Förp.
665-0394		1,679	$\pm 0,010$	25 °C	500 ml
665-0395		4,005	$\pm 0,010$	25 °C	500 ml
665-0396		6,865	$\pm 0,010$	25 °C	500 ml
665-0391		7,000	$\pm 0,010$	25 °C	500 ml
665-0397		7,413	$\pm 0,010$	25 °C	500 ml
665-0392		9,180	$\pm 0,010$		500 ml
665-0398		10,012	$\pm 0,010$	25 °C	500 ml
665-0399		12,45	$\pm 0,05$	25 °C	500 ml