

## 3M™ Attest™ Indicator 1262 Steam 100/bx



This 3M™ Attest™ Indicator is for 132 degrees C vacuum assisted or 121 degrees C gravity steam sterilizers. It provides assurance of sterility by challenging the sterilizer with quantifiable, highly resistant spores. Incubation is for 48 hours.