



OLYMPUS ENDODET, ENDOACT AND ENDODIS

PAA process chemistry

Olympus process chemicals for ETD machines provide comprehensive and safe reprocessing of your flexible, heat-sensitive endoscopes. With EndoDet, EndoAct and EndoDis, Olympus has created an optimally adapted, peracetic acid (PAA) based, process chemical system. Fully functional at low temperatures, it is largely effective while being fully compatible with your flexible Olympus endoscopes.

	EndoDet	EndoDis	EndoAct	
Product description	EndoDet is a surfactant-based detergent, which allows for optimum cleaning efficacy at lower temperature.	EndoDis is a disinfectant based on peracetic acid (PAA), providing a wide range of efficacy.	EndoAct allows the Olympus PAA process to work in a neutral pH value range and thus efficiently protects sensitive endoscope materials. It is an additive to EndoDis.	
			(I) Secretarian (I) Secretaria	
Composition	Non-ionic surfactants, glycols,	Active ingredient: Peroxyacetic acid	Phosphates, sodium hydroxide	
	solubilizers, cleaning booster	Other ingredients: Hydrogen peroxide		
		acetic acid, stabilizer		
Spectrum of	EndoDet effectively reduces organic	EndoDis, in combination with EndoAct is largely effective against relevant pathogens		
effect	debris and protein residues with special	in endoscopy and complies with European standards in state-of-the-art endoscope		
	cleaning booster.	reprocessing. Efficacy (under clean conditions):		
		EN 13727: Bactericidal activity		
		EN 13624: Fungicidal activity		
		EN 14348: Mycobactericidal activity	/	
		EN 14476: Virucidal activity		
		EN 13704: Sporicidal activity		
		Certificates and documentation available u	upon request	
Direction for use	EndoDet, EndoDis and EndoAct are all supplied as a concentrate. Place it into your ETD washer-disinfector according to the			
	instruction manual. The concentrate will then be automatically dosed into the machine process by the dosage pumps.			
Dosage (ETD3)	Dosage: 0.6% (6 ml/l)	Dosage: 1.2% (12 ml/l each)		
	Contact time: 3 min	Contact time: 5 min		
	Temperature: 35°C	Temperature: 35°C		
	According to standard ETD process settings			
Presentation	5 I canister	2.8 I canister	5 I canister	
	(sales unit: 3 canisters)	(sales unit: 3 canisters)	(sales unit: 3 canisters)	
Biocompatibility	Biocompatibility has been tested and proven according to ISO 10993. Simulated use tests demonstrated no influence on the viability of cells in cytotoxicity test from residues of the process chemicals			

Note: Protective equipment and clothing should always be worn when handling chemicals. Refer to the material safety data sheet.

Manufactured by: Ecolab Deutschland GmbH, Postfach 130406, D-40554 Düsseldorf, Tel:0211-9893-900, www.ecolab.com

